



EA MLA Signatory
Český institut pro akreditaci, o.p.s.
Olšanská 54/3, 130 00 Praha 3

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products, as amended

CERTIFICATE OF ACCREDITATION

No. 10/2022

JD Dvořák, s.r.o.
with registered office V Holešovičkách 1448/14, 180 00 Praha 8 - Libeň, Company Registration
No. 49621815

to the Calibration Laboratory No. 2298
JD Dvořák, s.r.o., Calibration Laboratory

Scope of accreditation:

Calibration of thermometers and hygrometers of thermal, climatic chambers and enclosures and measurement of temperature in temperature and humidity regulation equipment, single-point or multi-point time dependent measurement to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of Accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

ČSN EN ISO/IEC 17025:2018

In its activities performed within the scope and for the period of validity of this Certificate, the Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited Conformity Assessment Body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 196/2019 of 2. 5. 2019, or any administrative acts building upon it.

The Certificate of Accreditation is valid until: **2. 5. 2024**

Prague: 4. 1. 2022



Lukáš Burda
Director of the Department
of Testing and Calibration Laboratories
Czech Accreditation Institute
Public Service Company

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

JD Dvořák, s.r.o.

JD Dvořák, s.r.o., Calibration Laboratory
Tuřanka 1148/107, 627 00 Brno

Calibration laboratory locations:

1. JD Dvořák, s.r.o., Calibration Laboratory Tuřanka 1148/107, 627 00 Brno
2. JD Dvořák, s.r.o., Calibration Laboratory Toužimská 897/E3, 199 00 Praha 18

CMC for the field of measured quantity: Temperature

Ord. number ¹	Calibrated quantity / Subject of calibration	Nominal range				Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified ²	Calibration principle	Calibration procedure identification ³	Work-place
		min.	unit	max.	unit					
1*	Thermometers integrated in measuring chains of thermal and climatic chambers and enclosures and special-purpose enclosures, where heat can be generated	-70 °C	up to	-45 °C		0.20 °C	Comparison with a standard thermometer	Internal method I (DKD-R_5.7 method C)	1, 2	
		-45 °C	up to	100 °C		0.17 °C				
		100 °C	up to	200 °C		0.21 °C				
		200 °C	up to	300 °C		0.6 °C				
		300 °C	up to	400 °C		0.8 °C				
2*	Thermometers integrated in measuring chains of thermal and climatic chambers and enclosures and special-purpose enclosures, where heat can be generated	-70 °C	up to	-45 °C		0.35 °C	Comparison with a standard thermometer	Internal method 3 (DKD-R_5.7 method A and B)	1, 2	
		-45 °C	up to	100 °C		0.27 °C				
		100 °C	up to	150 °C		0.37 °C				
		150 °C	up to	180 °C		0.75 °C				

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

JD Dvořák, s.r.o.
JD Dvořák, s.r.o., Calibration Laboratory
Tuřanka 1148/107, 627 00 Brno

CMC for the field of measured quantity: Relative humidity

Ord. number ¹⁾	Calibrated quantity / Subject of calibration	Nominal range			Parameter(s) of the meas. quantity	Lowest expanded measurement uncertainty specified ²⁾	Calibration principle	Calibration procedure identification ³⁾	Work-place
		min. unit	max. unit	unit					
1*	Hygrometers integrated in measuring chains of climatic chambers and enclosures and special-purpose enclosures, where humidity and heat can be generated	10 % RH 65 % RH 90 % RH	up to up to up to	65 % RH 90 % RH 95 % RH		1.5 % RH 1.7 % RH 1.9 % RH	Comparison with a standard hygrometer	Internal method 2 (DKD-R_5.7 method C)	1, 2
2*	Hygrometers integrated in measuring chains of climatic chambers and enclosures and special-purpose enclosures, where humidity and heat can be generated	10 % RH 65 % RH 90 % RH	up to up to up to	65 % RH 90 % RH 95 % RH		1.5 % RH 1.7 % RH 1.9 % RH	Comparison with a standard hygrometer	Internal method 4 (DKD-R_5.7 method A and B)	1, 2

- 1) Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2) The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02 M, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- 3) If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).

